DNV-GL

EU-TYPE EXAMINATION CERTIFICATE

Certificate number 1075-17

Manufacturer Shenzhen Star Instrument Co., Ltd.

Contact person Mrs Li Wang

Address Star Industry Park, Baolong Industry City, Longgang District

Postal code, City 518116, Shenzhen

Country China

Production site Shenzhen Star Instrument Co., Ltd.

Address Star Industry Park, Baolong Industry City, Longgang District

Postal code, City 518116, Shenzhen

Country China

Instrument Electronic three-phase four-wire energy meter.

Direct connected

Mark - Type STK345, DTSY23S, DTSY23, DTS27, STC345, DTSY23, DTSYF23B, DTS27,

STZ345, DTSY23, DTS27

Register LCD

Accuracy Class Active: 1/C; Reactive: 1

Measurement range 100/173-127/220 V and 220/380-240/415 V; 0,15..5(100) A, 50 and 60 Hz

Active: 1000 imp./kWh; Reactive: 1000 imp./kvarh

Temperature range -40 .. 70 °C

Use Outdoor, not sensitive to phase sequence

Protection Class II

Environmental class M1, E1 and E2

Registry method Bi-directional method separate registers: received- and delivered energy of

the whole connection is added in separate registers

The energy meters meet the requirements of Directive 2014/32/EU, on the harmonisation of the laws of Member States relating to the making available on the market of measuring instruments (recast).

Certification was based on compliance with the following harmonised standards:

EN 50470-1 (2006) Electricity metering equipment (a.c.)-part 1: General requirements, tests

and test conditions - Metering equipment (class indexes A, B and C)

EN 50470-3 (2006) Electricity metering equipment (a.c.)-part 3: Particular requirements - Static

meters for active energy (class indexes A, B and C)

Valid until 13 July 2027

Based on a non-recurrent examination.

Version 1 of this certificate. The results are recorded in the following annex: test report 1290-17.

DNV GL Netherlands B.V.

J.P. Eonteijne

Executive Vice President KEMA Laboratories

Arnhem, 13 July 2017



Copyright: Only integral reproduction of this Certificate is permitted without written permission from DNV GL.