

## EU-TYPE EXAMINATION CERTIFICATE

Certificate number	1389-16
Manufacturer	ZIV Metering Solutions S.L.
Contact person	Ms. Cristina Martinez
Address	Parque Tecnológico, 210
Postal code, Place	48170 Zamudio
Country	Spain
Production site	ZIV Metering Solutions S.L.
Address	Parque Tecnológico, 210
Postal code, Place	48170 Zamudio
Country	Spain
Instrument	Electronic single-phase two-wire energy meter Direct connected
Mark - Type	5CTME2C477
Register	LCD
Accuracy Class	1 / B for active energy, 2 for reactive energy
Measurement range	127 - 230 V 0,25..5(80) A 50 Hz Active: 1000 imp./kWh; Reactive: 1000 imp./kvarh
Temperature range	-25..70 °C
Use	Indoor
Protection Class	II
Environmental class	M1, E1
Registry method	Bi-directional method separate registers: received- and delivered energy of the whole connection is added in separate registers

The energy meter meets the requirements of Directive 2014/32/EU of the European parliament and the council of 26 February 2014 on measuring instruments.


Certification was based on compliance with the following harmonised standards:

EN 50470-1 (2006)	Electricity metering equipment (a.c.)-part 1: General requirements, tests and test conditions - Metering equipment (class indexes A, B and C)
EN 50470-3 (2006)	Electricity metering equipment (a.c.)-part 3: Particular requirements - Static meters for active energy (class indexes A, B and C)
Valid until	1 September 2026

Based on a non-recurrent examination.

Version 1 of this certificate. The results are recorded in the following annex: test report 1387-16.

KEMA Nederland B.V.



J.P. Fonteijne  
Executive Vice President  
KEMA Laboratories



Arnhem, 1 September 2016

Copyright: Only integral reproduction of this Certificate is permitted without written permission from DNV GL.

KEMA Nederland B.V. Utrechtseweg 310, 6812 AR Arnhem, P.O. Box 9035, 6800 ET Arnhem, the Netherlands  
Notified body number 2290