



EC-type examination certificate

Number **T10026** Revision 0
Projectnumber: 511156
Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer ENERMET OY
Salvesenintie 6
FIN-40420 JYSKÄ
Finland

In respect of A model of a static **Active Electrical Energy Meter**

Type : E130-(t)(s)

Characteristics Reference voltage : 230 V
Reference current : 5 or 10 A
Destined for the measurement of : electrical energy, in a single-phase two-wire network
Accuracy class : A or B
Mechanical environment class : M1
Electromagnetic environment class : E2
Temperature range : -40°C / +70°C

In the description number T10029 revision 0 further characteristics are described.

Valid until 22 December 2016

Description and documentation The instrument is described in the description number T10029 revision 0 and documented in the documentation folder T10029-1, appertaining to this EC-type examination certificate.

Dordrecht, 22 December 2006
NMI Certin B.V.

bla H.S. Schout

Ing. C. Oosterman
Manager Product Certification

