

Issued by NMI Certin B.V.,
 designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:

Manufacturer Itron
 Langer Road, Felixstowe
 Suffolk IP11 2ER
 England

Measuring instrument **A static Active Electrical Energy Meter**

Type : ACE3000 type 100 and 110 (drum register) or type 420 or type 520 (LCD)

Reference voltage : 3 x 120/208 V3 x 240/415 V (see 1.2.3)

Reference current : 5, 10 or 20 A

Destined for the measurement of : electrical energy, in a
 - three-phase four-wire network
 - three-phase three-wire network
 - two-phase three-wire network
 - single-phase two-wire network

Accuracy class : A or B

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes
 – Description T10050 revision 13
 – Documentation folder T10050-8

Valid until 2 February 2017

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
 10 July 2012

C. Oosterman
 Head Certification Board