



Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Landis+Gyr S.p.A.
Via del Maggiolino, 53/57
00155 Roma
Italy

In respect of A model of a static **Active Electrical Energy Meter**

Type : LENNT-P1, LENNT-P1-S, LENNT-P3
LENNT-P1A, LENNT-P1A-S, LENNT-P3A

Characteristics Reference voltage : 3 x 230/400 V
Reference current : 5 A
Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
Accuracy class : B
Mechanical environment class : M2
Electromagnetic environment class : E2
Temperature range : -40°C / +70°C

In the description number T10130 revision 1 further characteristics are described.

Valid until 24 April 2018

Description and documentation The instrument is described in the description number T10130 revision 1 and documented in the documentation folder T10130-2, appertaining to this EC-type examination certificate.

Remark This version replaces the earlier version, including its documentation folder.

Dordrecht, 6 October 2008
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification