



Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Actaris  
Langer Road, Felixstowe  
Suffolk IP11 2ER  
England

In respect of A model of a static **Active Electrical Energy Meter**

Type : ACE4000 PSS

Characteristics Reference voltage : 230 V  
Reference current : 10 A  
Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network  
Accuracy class : B or A  
Mechanical environment class : M1  
Electromagnetic environment class : E2  
Temperature range : -40°C / +70°C

In the description number T10160 revision 3 further characteristics are described.

Valid until 19 November 2018

Description and documentation The instrument is described in the description number T10160 revision 3 and documented in the documentation folder T10160-2, appertaining to this EC-type examination certificate.

Remark This version replaces the earlier version, except for its documentation folder.

Dordrecht, 2 December 2008  
NMI Certin B.V.

  
Ing. C. Oosterman  
Manager Product Certification