

- Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
- Manufacturer SAMGAS S.r.l.
Strad Provinciale 33, I-20080 Vernate, Italy
- Measuring instrument **A Diaphragm Gas Meter**
- Type : RS / 2001 LA (steel housing)
RS / 2001 AL (aluminium housing)
RS / 2,4 (steel housing)
- Destined for the measurement of : Gas volume
- Accuracy class : Class 1,5
- Environment classes : M2 / E1
- Location : Closed
- Temperature range : - 10 °C / +40 °C

G-value	Q _{max} [m ³ /h]	Q _{min} [m ³ /h]	Q _t [m ³ /h]	V [dm ³]
6	10	0,060	1,0	2,4
4	6	0,040	0,6	2,4
4	6	0,040	0,6	1,2
2,5	4	0,025	0,4	1,2
1,6	2,5	0,016	0,25	1,2

Further properties are described in the annexes

- Description T10226 revision 3
- Documentation folder T10226-3

- Valid until 16 September 2019
- Remarks This revision replaces the earlier versions, including its documentation folder.

- Issuing Authority **NMI Certin B.V., Notified Body number 0122**
25 June 2012

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).

Reproduction of the complete document only is permitted.