

**Issued by** NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 9 of Directive 2004/22/EC, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2004/22/EC, to:

**Manufacturer** Itron France  
1 Avenue des temps modernes  
ZI du Bernais B.P. 23  
86361 Chasseneuil du Poitou  
France

**Measuring instrument** A static **Active Electrical Energy Meter**  
Type : ACE431E  
Reference voltage : 230V  
Reference current : 15 A  
Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network  
Accuracy class : A or B  
Environment classes : M1 / E2  
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes  
– Description T10259 revision 1  
– Documentation folder T10259-1

**Valid until** 9 February 2020

**Remark** This version replaces the earlier version, except for its documentation folder.

**Issuing Authority** **NMI Certin B.V., Notified Body number 0122**  
17 February 2010

  
C. Oosterman  
Head of the Certification Board