

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Landis + Gyr
1 Lysander Drive, Northfields Industrial Estate
Market Deeping
Peterborough PE6 8FB
England

Measuring instrument

A static **Active Electrical Energy Meter**

Type : E450 IDIS ZMXi 3x0 xP xx xx xx . xx S2
E450 IDIS ZFXi 3x0 xP xx xx xx . xx S2
E450 Non-IDIS ZMX 3x0 xP xx xx xx.xx
S2

Manufacturer's mark or name : Landis + Gyr

Reference voltage : 3x230/400 V or 3x230 V or 230 V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- three-phase three-wire network
- two-phase three-wire network
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T10471 revision 8;
- Documentation folder T10471-3.

Valid until

6 July 2022

Remarks

This revision replaces the earlier versions, except its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
29 July 2019


C. Oosterman
Head Certification Board