

## EU type-examination certificate

Number **T10576** revision 3 Project number 1900879 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable

requirements of Directive 2014/32/EU, to:

Manufacturer AEM S.A.

Calea Buziasului 26 300693 TIMISOARA

Romania

Measuring instrument A static Active Electrical Energy Meter

Type : CSM 0203

Reference voltage : 230 V Reference current : 5 A

Destined for the measurement of : electrical energy, in a

single-phase two-wire network

Accuracy class : A or B Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

Description T10576 revision 3;Documentation folder T10576-3.

Valid until 4 4 4 November 2023

Remarks This revision replaces the earlier versions, including its documentation

folder.

Issuing Authority NMi Certin B.V., Notified Body number 0122

17 July 2017

C. Oosterman

Head Certification Board

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Reproduction of the complete document only is permitted.

