

EC type-examination Certificate

Number **T10589** revision 0 Project number 13200396 Page 1 of 1

Issued by Manufacturer	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to: Shenzhen Kaifa Technology Co., Ltd. 7006 Caitian RD.,Futian Dist. 518035 Shenzhen												
	P.R. China												
Measuring instrument	· · · · · · · · · · · · · · · · · · ·												
	Type + + + + + + + + +	÷÷.	MA301										
	Reference voltage Reference current Destined for the measurement of	+ + + + + + + +	3x230/400 V 5 A electrical energy, in a - three-phase four-wire network										
		+ +	+ + +										
	Accuracy class Environment classes Temperature range	+ + + + + +	A or B M2 / E2 -40 °C /		°C								
	Further properties are described in – Description T10589 revision 0; – Documentation folder T10589-1.	the a	nnexes:										
Valid until	12 December 2023												
+ + + + + + + + +													
+ $+$ $+$ $+$ $+$ $+$ $+$ $+$ $+$ $+$		+ +	+ + +	+ +									
Issuing Authority	NMi certin B.V., Notified Body 12 December 2013	amun	er 0122	1.1									
	11 Luci												
	+ + + + + + + + + + + + + + + + + + + +												
• • • • • • • • · · Z	C. Oosterman Head Certification Board												
NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl	This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability. The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nando/	Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl). Reproduction of the complete document only is permitted.								YECTI	ON 22		