

EU-type examination certificate

Number **T10756** revision 5 Project number 1901846 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable

requirements of Directive 2014/32/EU, to:

Manufacturer Elster Solutions SAS

1 allée de Bruxelles

Estrées Deniecourt (80200)

France

Authorised Elster Solutions GmbH representative Steinern Str. 19-21

55252 Mainz-Kastel Hessen

Germany

Measuring instrument A static Active Electrical Energy Meter

Type : AS330L-G3

Manufacturer's mark or name 🔹 🔹 : Elster or Honeywell

Reference voltage : 230 V Reference current : 10 A

Destined for the measurement of : electrical energy, in a

- single-phase two-wire network

Accuracy class : E

Environment classes : M1 / E2

Temperature range \div : $-40 \,^{\circ}\text{C} / +70 \,^{\circ}\text{C}$

Further properties are described in the annexes:

Description T10756 revision 5;Documentation folder T10756-2.

Valid until 10 February 2026

Remark This revision replaces the previous versions, except for its documentation

folder

Issuing Authority NMi Certin B.V., Notified Body number 0122

17 January 2018

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1

3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at

http://ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete document only is permitted.

