

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Shenzhen Kaifa Technology (Chengdu) Co., Ltd. No. 99 Tianquan Rd. Hi-Tech Development Zone Chengdu, (PC: 611731) P.R. China
Measuring instrument	A static <b>Active Electrical Energy Meter</b> Type : MA304H3, MA304H4, MA304H3E, MA304H4D as direct connected meter (DCM) MA304T3, MA304T4 as transformer connected meter (TCM)
	Manufacturer's mark or name : KAIFA
	Reference voltage : 3x230/400 V or 3x230 V
	Reference current : 5 A (DCM) 1 A (TCM)
	Destined for the measurement of : electrical energy, in a - three-phase four-wire network - two-phase three-wire network
	Accuracy class : B or A (for DCM) C or B (for TCM)
	Environment classes : M2 / E2
	Temperature range : -40 °C / +70 °C
	Further properties are described in the annexes: – Description T10806 revision 7; – Documentation folder T10806-6.
Valid until	25 August 2026
Remarks	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
15 March 2022

## Certification Board

**NMi Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.