


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Rayleigh Instruments Limited Raytel House, Cutlers Road South Woodham Ferrers Chelmsford, Essex CM3 5WA United Kingdom
Measuring instrument	A static Active Electrical Energy Meter Type : RI-70-100-A, RI-70-100-P, RI-D70-C and RI-D70-MB Manufacturer's mark or name : Rayleigh Instruments Reference voltage : 3x230/400 V Reference current : 5 or 10 A Destined for the measurement of : electrical energy, in a - three-phase four-wire network Accuracy class : A or B Environment classes : M1 / E2 Temperature range : -25 °C / +55 °C Further properties are described in the annexes: - Description T11065 revision 1; - Documentation folder T11065-2.
Valid until	14 April 2027
Remark	This revision replaces the earlier version, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
28 May 2018



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.