

EU-type examination certificate

Number **T11080** revision 0 Project number 1900989 Page 1 of 1

Issued by + + + + NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable

requirements of Directive 2014/32/EU, to:

Manufacturer Owen Brothers Metering UK Ltd.

Glen Trading Estate, Wellyhole Street OL4 3BF Oldham, Greater Manchester

UK

Measuring instrument A static Active Electrical Energy Meter

Type : OB111, OB112

Manufacturer's mark or name : Owen Brothers

Reference voltage : 230 V Reference current : 5 A

Destined for the measurement of : electrical energy, in a

- single-phase two-wire network

Accuracy class : A or B Environment classes : M1 / E2

Temperature range : -25 °C/+55 °C

Further properties are described in the annexes:

Description T11080 revision 0;Documentation folder T11080-1.

Valid until 26 April 2027

Issuing Authority NMi Certin B.V., Notified Body number 0122

26 April 2017

Costerman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Reproduction of the complete document only is permitted.

