

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer Klefr  
Pondweg 7  
2153 PK Nieuw Vennep  
The Netherlands

Measuring instrument A static **Active Electrical Energy Meter**

Type : 6941, 6943, 6944

Manufacturer's mark or name : Klefr

Reference voltage : 3x230/400 V

Reference current : 1,5 A

Destined for the measurement of : electrical energy, in a  
- three-phase four-wire network  
- three-phase three-wire network  
- single-phase two-wire network

Accuracy class : C

Environment classes : M1 / E2

Temperature range : -25 °C / +70 °C

Further properties are described in the annexes:

- Description T11176 revision 1;
- Documentation folder T11176-2.

Valid until 1 September 2027

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority

**NMI Certin B.V., Notified Body number 0122**  
25 June 2021

Certification Board

**NMI Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.