

EU-type examination certificate



Number **T11252** revision 7 Project number 3488883 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable

requirements of Directive 2014/32/EU, to:

Manufacturer Shenzhen Kaifa Technology (Chengdu) Co., Ltd.

No. 99 Tianquan Rd.,

Hi-Tech Development Zone,

Chengdu,

P.R.C. (PC.611731)

Measuring instrument A static Active Electrical Energy Meter

Type : MB310H4BDE, MB310H4BDE1,

MB310H4BNL and MB310H4BNL5

Manufacturer's mark or name : Kaifa

Reference voltage : 3x230/400 V Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a

three-phase four-wire networksingle-phase two-wire network

Accuracy class : B or A Environment classes : M1 / E2

Temperature range : $-40 \,^{\circ}\text{C} / +70 \,^{\circ}\text{C}$

Further properties are described in the annexes:

Description T11252 revision 7;Documentation folder T11252-5.

Valid until 26 March 2028

Remark This revision replaces the previous versions, including its documentation

folder.

Issuing Authority NMi Certin B.V., Notified Body number 0122

29 April 2022

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http://ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



