

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 17 of Directive 2014/32/EU, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer ZTE Corporation
NO.55 Hi-Tech Road South
NanShan District
Shenzhen City, Guangdong Province, P.R.China

Measuring instrument A static **Active Electrical Energy Meter**

Type	: ZX AE230
Manufacturer's mark or name	: ZTE Corporation
Reference voltage	: 3x57.7V/100V, 3x230V/400V, 3x100V, 3x240V/415V
Reference current	: 1A
Destined for the measurement of	: electrical energy, in a - three-phase four-wire network - three-phase three-wire network
Accuracy class	: B or C
Environment classes	: M1 / E2
Temperature range	: -40 °C / +70 °C

Further properties are described in the annexes
– Description T11337 revision 0
– Documentation folder T11337-1

Valid until 19 July 2029

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
19 July 2019


C. Oosterman
Head Certification Board