

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer ITRON
1 Avenue des temps modernes
ZI du Bernais B.P. 23
86361 Chasseneuil du Poitou
France

Measuring instrument A static **Active Electrical Energy Meter**

Type	:	ITE411L3N2
Manufacturer's mark or name	:	Itron
Reference voltage	:	230 V
Reference current	:	10 A
Destined for the measurement of	:	electrical energy, in a - single-phase two-wire network
Accuracy class	:	A or B
Environment classes	:	M1 / E2
Temperature range	:	-25 °C / +55 °C

Further properties are described in the annexes:
– Description T11444 revision 3;
– Documentation folder T11444-4.

Valid until 9 November 2028

Initially issued 9 November 2018

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
20 October 2023

Certification Board