

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer ITRON  
1 Avenue des temps modernes  
ZI du Bernais B.P. 23  
86361 Chasseneuil du Poitou  
France

Measuring instrument A static **Active Electrical Energy Meter**

Type : ITE411L3B2  
Manufacturer's mark or name : Itron  
Reference voltage : 230 V  
Reference current : 10 A  
Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network  
Accuracy class : A or B  
Environment classes : M1 / E2  
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:  
- Description T11460 revision 4;  
- Documentation folder T11460-4.

Valid until 9 November 2028

Initially issued 9 November 2018

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
20 October 2023

Certification Board