

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Conrad Electronic SE
Klaus-Conrad-Str. 1
D-92240 Hirschau
Germany

Measuring instrument A static **Active Electrical Energy Meter**

Type	:	DPM-1200 C-CONTROL
Manufacturer's mark or name	:	Conrad Electronic SE
Reference voltage	:	230 V
Reference current	:	5 A
Destined for the measurement of	:	electrical energy, in a - single-phase two-wire network
Accuracy class	:	A or B
Environment classes	:	M1 / E2
Temperature range	:	-25 °C / +55 °C

- Further properties are described in the annexes:
- Description T11538 revision 0;
 - Documentation folder T11538-1.

Valid until 28 March 2029

Issuing Authority

NMI Certin B.V., Notified Body number 0122
28 March 2019


C. Oosterman
Head Certification Board