

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer KŘIŽÍK s.r.o.
Lobezská 214/9
326 00 Plzeň
Czech Republic

Measuring instrument A static **Active Electrical Energy Meter**

Type : H100

Manufacturer's mark or name : KŘIŽÍK s.r.o.

Reference voltage : 240 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : B

Environment classes : M1 / E2

Temperature range : -25 °C to +70 °C

Further properties are described in the annexes:
– Description T12381 revision 1;
– Documentation folder T12381-1.

Valid until 18 July 2032

Remark This revision replaces the previous version(s), except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
5 December 2022

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.