



EU-type examination certificate

Number **T10358** revision 25
Project number 4112488
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets the
applicable requirements of Directive 2014/32/EU, to:

Applicant

ABB S.p.A.
Via Vittor Pisani 16
20124 Milano
Italy

Manufacturer

ABB S.p.A.
Viale dell'Industria 18
20009 Vittuone (MI)
Italy

Measuring instrument

An **Active Electrical Energy Meter**

Type : A41
A42

Manufacturer's mark or name : ABB

Reference voltage : 57,7...288 V, 100...288 V (A42...-12.)

Reference current : A41: 5 A (direct connected)
A42: 1 A (indirect connected)

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : A41: A or B
A42: A, B or C

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description 10358 revision 25;
- Documentation folder 10358-13.

Initially issued

26 January 2012

Valid until

26 January 2032

Remark

This revision replaces the earlier versions, including for its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
18 June 2026

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The appointment of NMi Certin UK (TIC) Ltd.,
as Approved Body can be verified at
<https://www.gov.uk/uk-market-conformity-assessment-bodies>

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

