



EU-type examination certificate

Number **T10916** revision 21
Project number 3988487
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets the
applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Networked Energy Services
780 Montague Expressway, Suite 401
San Jose, California 95131
United States of America

Measuring
instrument

A static Active Electrical Energy Meter

Type

: Direct Connected Meter (DCM)
83334-3xxxx, 84334-3xxxx, 85334-3xxxx
Transformer Connected Meter (TCM)
83534-3xxxx, 84534-3xxxx, 85534-3xxxx
83524-3xxxx, 84524-3xxxx, 85524-3xxxx

Manufacturer's mark or name : Networked Energy Services

Reference voltage : 3x110/190 V ... 3x230/400 V
2x110 V ... 2x230 V
110 V ... 230 V

Reference current : 5 A (DCM/TCM), 10 A and 20 A (DCM)

Destined for the
measurement of : electrical energy, in a
- three-phase four-wire network
- two-phase three-wire network
- single-phase two-wire network

Accuracy class : A, B, C

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T10916 revision 21;
- Documentation folder T10916-13.

Valid until

4 April 2027

Initially issued

4 April 2017

Remark

This revision replaces the earlier versions, including its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
8 January 2026

Certification Board

NMi Certin B.V.
Trijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the manufacturer
shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.


PRODUCTS
RvA C 681