





Number **T10995** revision 10 Project number 3777290 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity asssessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets

the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Shenzhen Kaifa Technology (Chengdu) Co., Ltd.

No. 99 Tianquan Rd.

Hi-Tech Development Zone

Chengdu

P.R. China (PC.611731)

Measuring instrument A static Active Electrical Energy Meter

Type : MA109P

Manufacturer's mark or name : KAIFA

Reference voltage : 230 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a

- single-phase two-wire network

Accuracy class : B

Environment classes : M1 / E2

Temperature range : $-40 \,^{\circ}\text{C} / +70 \,^{\circ}\text{C}$

Further properties are described in the annexes:

- Description T10995 revision 10;

- Documentation folder T10995-7.

Valid until 12 May 2027

Initially issued 12 May 2017

This revision replaces the earlier versions, except for its documentation

folder.

Issuing Authority NMi Certin B.V., Notified Body number 0122

19 March 2024

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

Remarks

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified

Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/ Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



