

Issued by NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer MetCom Solutions GmbH
 Marie-Curie-Strasse 19
 D-68219 Mannheim
 Germany

Measuring instrument An **electrical energy meter**

Type : MCS301 and MCS501

Manufacturer's mark or name : MetCom

Reference voltage : 3x58/100 V, 3x230/400 V or 3x58/100 V 3x240/415 V or 3x58/100 V 3x277/480 V or 3x100 V ... 3x240 V or 3x230V/400V ... 3x415/720V 3x400/690V (Aux Power)

Reference current : 1 A or 5 A (CT version) 5 A or 10 A (DC version)

Destined for the measurement of : electrical energy, in a
 - three-phase four-wire network
 - three-phase three-wire network
 - two-phase three-wire network
 - single-phase two-wire network

Accuracy class : A or B (DC version) B or C (CT version)

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11028 revision 15;
- Documentation folder T11028-12.

Valid until 16 October 2027

Initially issued 16 October 2017

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
 23 May 2024

Certification Board