

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis + Gyr A.E.  
78th km National Road Athens-Corinth  
GR-201 00, Corinth  
Greece

Measuring instrument A static **Active Electrical Energy Meter**

Type : E450 G3 IDIS (ZCXi1x0xQ S4)  
Manufacturer's mark or name : Landis + Gyr  
Reference voltage : 230 V  
Reference current : 5 A or 10 A  
Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network  
Accuracy class : B or A  
Environment classes : M2 / E2  
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:  
- Description T11169 revision 12;  
- Documentation folder T11169-3.

Valid until 1 September 2027

Initially issued 1 September 2017

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
16 December 2024

Certification Board