

EU-type examination certificate

Number **T11234** revision 4
Project number 4010733
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Inepro Metering BV
Pondweg 7
2153 PK Nieuw Vennep
The Netherlands

Measuring instrument

A static **Active Electrical Energy Meter**

Type : 0318

Manufacturer's mark or name : Inepro

Reference voltage : 3x230/400 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11234 revision 4;
- Documentation folder T11234-4;
- Evaluation report ER11234 revision 4.

Valid until

28 May 2028

Initially issued

28 May 2018

Remark

This revision replaces the earlier versions, including its documentation
folder.

Issuing Authority

NMi Certin B.V. Notified Body number 0122
27 August 2026

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the manufacturer
shall indemnify third-party liability.

The appointment of NMi Certin UK (TIC) Ltd., as
Approved Body can be verified at
[https://www.gov.uk/uk-market-conformity-
assessment-bodies](https://www.gov.uk/uk-market-conformity-assessment-bodies)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon
at the top of the electronic
version of this certificate.