

Issued by NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr AG  
Alte Steinhauserstrasse 18  
CH-6330 Cham  
Switzerland

Measuring instrument A static **Active Electrical Energy Meter**

Type : E570 S2 ZxY4xxCQ...S2 (G3-PLC) or  
ZxY4xxCW1...S2 (P2P 2G/4G)

Manufacturer's mark or name : Landis+Gyr

Reference voltage : ZxY4xxCQ...S2 (G3-PLC)  
ZMY 3x220/380V – 3x240/415V  
ZFY 3x220V – 3x240V  
ZxY4xxCW1...S2 (P2P 2G/4G)  
ZMY 3x58/100V - 3x277/480V  
ZFY 3x100V – 3x240V

Reference current : 1 A or 5 A

Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network  
- two-phase three-wire network  
- three-phase three-wire network  
- three-phase four-wire network

Accuracy class : B or C

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11335 revision 11;
- Documentation folder T11335-5.

Initially issued 8 August 2018

Valid until 8 August 2028

Remark This revision replaces the earlier versions, excluding its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
7 March 2024

Certification Board