

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Gridspertise S.r.L  
Via Ombrone, 2  
00198 Roma  
Italia

Measuring instrument A static **Active Electrical Energy Meter**  
Type : GESIS  
Manufacturer's mark or name : Gridspertise  
Reference voltage : 3x230/400 V  
3x127/220 ... 3x230/400 V  
Reference current : 1 A  
Destined for the measurement of : electrical energy, in a  
- three-phase four-wire network  
Accuracy class : B or A  
Environment classes : M1 / E2  
Temperature range : -25 °C / +55 °C

Further properties are described in the annexes;  
- Description T11471 revision 24;  
- Documentation folder T11471-8.

Valid until 19 December 2028

Initially issued 19 December 2018

Remark This revision replaces the previous versions, except its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
05 June 2024

Certification Board