

Issued by NMI Certin B.V.,  
 designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr AG  
 Alte Steinhäuserstrasse 18  
 CH-6330 Cham  
 Switzerland

Measuring instrument A static **Active Electrical Energy Meter**

- Type : E660-A...S1 and E660-A...S2
- Manufacturer's mark or name : Landis+Gyr
- Reference voltage : 3x58/100V - 3x240/415V - 3x277/480V (3p/4w)  
 3x100V – 3x480V (3p/3w)
- Reference current : 1 or 5 A
- Destined for the measurement of : electrical energy, in a
  - three-phase four-wire network
  - three-phase three-wire network
  - two-phase three-wire network
  - single-phase two-wire network
- Accuracy class : B or C
- Environment classes : M2 / E2
- Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11845 revision 11;
- Documentation folder T11845-6.

Valid until 18 June 2030

Initially issued 18 June 2020

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
 29 October 2024

### Certification Board