

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Networked Energy Services
780 Montague Expressway, Suite 401
San Jose, California 95131
United States of America

Measuring instrument A static **Active Electrical Energy Meter**
Type : Direct Connected Meter (DCM)
83335-3xxxxxx, 85335-3xxxxxx
Transformer Connected Meter (TCM)
83535-3xxxxxx, 85535-3xxxxxx

Reference voltage : 3x110/190 V ... 3x240/415 V
2x110 V ... 2x240 V
110 V ... 240 V

Reference current : 5 A (TCM)
5 A, 10A, 20 A (DCM)

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- two-phase three-wire network
- single-phase two-wire network

Accuracy class : A or B (DCM)
B or C (TCM)

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12242 revision 1;
- Documentation folder T12242-1.

Valid until 14 March 2033

Initially issued 14 March 2023

Remarks This revision replaces the earlier version, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
18 November 2024

Certification Board