



EU-type examination certificate

Number **T12242** revision 3
Project number 3988487
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Networked Energy Services
780 Montague Expressway, Suite 401
San Jose, California 95131
United States of America

Measuring instrument

A static **Active Electrical Energy Meter**

Type

: Direct Connected Meter (DCM)
83335-3xxxxxx, 85335-3xxxxxx
Transformer Connected Meter (TCM)
83535-3xxxxxx, 85535-3xxxxxx,
83525-3xxxxxx, 85525-3xxxxxx

Reference voltage

: 3x110/190 V ... 3x240/415 V
2x110 V ... 2x240 V
110 V ... 240 V

Reference current

: 5 A (TCM)
5 A, 10A, 20 A (DCM)

Destined for the
measurement of

: electrical energy, in a
- three-phase four-wire network
- two-phase three-wire network
- single-phase two-wire network

Accuracy class

: A or B (DCM)
B or C (TCM)

Environment classes

: M1 / E2

Temperature range

: -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12242 revision 3;
- Documentation folder T12242-3.

Valid until

14 March 2033

Initially issued

14 March 2023

Remarks

This revision replaces the earlier version, including its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
8 January 2026

Certification Board

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

