

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Zhejiang Eastron Electronic Co., Ltd.
No.52, Dongjin Road, Nanhu
Jiaxing, Zhejiang
China

Measuring instrument A static **Active Electrical Energy Meter**

Type	: SDM230
Manufacturer's mark or name	: Eastron
Reference voltage	: 230 V
Reference current	: 10 A
Destined for the measurement of	: electrical energy, in a - single-phase two-wire network
Accuracy class	: B or C
Environmental classes	: M1 / E2
Temperature range	: -40 °C / +70 °C

Further properties are described in the annexes
– Description T12800 revision 1
– Documentation folder T12800-2

Valid until 2 December 2034

Initially issued 2 December 2024

Remarks This revision replaces the earlier version, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
26 May 2026

Certification Board