

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Applied Micro Electronics "AME" B.V.
Esp 100
5633 AA Eindhoven
Netherlands

Measuring instrument **A static Active Electrical Energy Meter integrated in an EV Charging System**

Type : AACP
Manufacturer's mark or name : AME
Reference voltage : 3x230/400 V or 230 V
Reference current : 5 A
Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- single-phase two-wire network
Accuracy class : A or B
Environment classes : M1 / E2
Temperature range : -25 °C / +55 °C

Further properties are described in the annexes:
- Description T12803 revision 0;
- Documentation folder T12803-1.

Valid until 20 June 2034

Initially issued 20 June 2024

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
20 June 2024

Certification Board