



EU-type examination certificate

Number **T12814** revision 2
Project number 4033215
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Enphase Energy Inc.
47281 Bayside Parkway
Fremont, CA 94538
United States of America

Authorised
representative

Enphase Energy NL B.V.
Het Zuiderkruis 65
5215 MV 's-Hertogenbosch
The Netherlands

Measuring instrument

A static **Active Electrical Energy Meter integrated in an EV Charging
System**

Type : IQ EV Charger 2

Manufacturer's mark or name : Enphase Energy Inc

Reference voltage : 3x230/400 V, 3x230 V, 230 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- three-phase three-wire network
- single-phase two-wire network

Accuracy class : B or A

Environment classes : M1 / E1

Temperature range : -40 °C / +55 °C

Further properties are described in the annexes

- Description T12814 revision 2
- Documentation folder T12814-2

Valid until

18 November 2034

Initially issued

18 November 2024

Remarks

This revision replaces the earlier versions, including its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
7 November 2025

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

