





Number T12891 revision 0 Project number 3859650 Page 1 of 1

Issued by NMi Certin B.V.,

> designated and notified by the Netherlands to perform tasks with respect to conformity asssessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets

the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Kaifa Technology Netherlands B.V.

Burgemeester Burgerslaan 40

5245 NH Rosmalen The Netherlands

Measuring instrument A static Active Electrical Energy Meter

Type MA110M

Manufacturer's mark or name Kaifa Technology Netherlands B.V.

Reference voltage 230 or 240 V

Reference current 5 A

Destined for the measurement of electrical energy, in a

- single-phase two-wire network

Accuracy class В :

M1 / E2 or M2 / E2 **Environment classes**

Temperature range -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12891 revision 0;

- Documentation folder T12891-1.

Valid until 30 August 2034

Initially issued 30 August 2024





NMi Certin B.V., Notified Body number 0122 30 August 2024

Certification Board

This document is issued under the provision that no liability is accepted and that the

manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete document only is permitted.

and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl

