

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Itron Indonesia - PT Mecoindo  
EJIP Plot 6B-2 - Cikarang Selatan Bekasi  
17550 Jawa Barat  
Indonesia

Measuring instrument A static **Active Electrical Energy Meter**

Type : Riva T  
Manufacturer's mark or name : Itron Indonesia – PT Mecoindo  
Reference voltage : 3x220/380 , 3x230/400 , 3x240/415  
Reference current : 5 A  
Destined for the measurement of : electrical energy, in a  
- three-phase four-wire network  
-three-phase three-wire network

Accuracy class : A or B  
Environment classes : M1 / E2  
Temperature range : -40 °C / +55 °C

Further properties are described in the annexes:  
– Description T12905 revision 2;  
– Documentation folder T12905-3.

Valid until 14 August 2035

Initially issued 14 August 2025

Remark This revision replaces the earlier version, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
6 March 2026

Certification Board