

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Gridspertise Srl
Via Ombrone 2
00198 Roma
Italia

Measuring instrument A static **Active Electrical Energy Meter**

Type : Gridspertise
Manufacturer's mark or name : Globy-T
Reference voltage : 3x230/400 V; 3x240/415 V
Reference current : 5 A
Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
Accuracy class : B
Environment classes : M1 / E2
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:
- Description T12909 revision 0;
- Documentation folder T12909-1.

Valid until 14 February 2035

Initially issued 14 February 2025

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
14 February 2025

Certification Board