

EU-type examination certificate

Number **T13062** revision 0
Project number 3979460
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Rayleigh Instruments Limited.
Raytel House, 1-5 Cutlers Road
South Woodham Ferrers
Chelmsford, Essex CM3 5WA
United Kingdom

Measuring instrument A static **Active Electrical Energy Meter**

Type : RI-70-100-A, RI-D70-A, RI-70-100-P,
RI-D70-P, RI-D70-B, RI-D70-R, RI-D70-C,
RI-D70-MB, RI-D70-C, RI-D70-MB,
RI-D70CT-P, RI-D70CT-B, RI-D70CT-R,
RI-D70CT-C, RI-D70CT-MB

Manufacturer's mark or name : Rayleigh Instruments Limited.

Reference voltage : 3x230/400 V

Reference current : 5 or 10 A (RI-70-100-A, RI-D70-A, RI-70-
100-P, RI-D70-P, RI-D70-B, RI-D70-R, RI-
D70-C, RI-D70-MB)
10 A (RI-D70-C, RI-D70-MB)
1,5 A (RI-D70CT-P, RI-D70CT-B, RI-
D70CT-R, RI-D70CT-C, RI-D70CT-MB)

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T13062 revision 0;
- Documentation folder T13062 -1.

Valid until 10 July 2035

Initially issued 10 July 2025

Issuing Authority

NMi Certin B.V., Notified Body number 0122
10 July 2025

Certification Board