

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Itron Indonesia - PT Mecoindo  
EJIP Plot 6B-2 - Cikarang Selatan Bekasi  
17550 Jawa Barat  
Indonesia

Measuring instrument A static **Active Electrical Energy Meter**

Type : RIVA S

Manufacturer's mark or name : Itron

Reference voltage : 220...240 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +55 °C

Further properties are described in the annexes:  
- Description T13079 revision 1;  
- Documentation folder T13079-1.

Valid until 12 February 2036

Initially issued 12 February 2026

Remark This revision replaces the earlier version(s), except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
26 February 2026

Certification Board