

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Accuenergy
440 Comstock Road
Toronto, Ontario, M1L2H6
Canada

Measuring instrument **A static DC electrical energy meter** intended to be integrated in a EV charging system

Type : AcuDC-301
Manufacturer's mark or name : Accuenergy
Reference voltage : 60 VDC ... 1000 VDC
Reference current : 130 A
Destined for the measurement of : electrical energy, in a
- DC electrical network
Accuracy class : C
Environment classes : M1 / E2
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:
- Description T13121 revision 1;
- Documentation folder T13121-2.

Valid until 16 February 2036

Initially issued 16 February 2026

Remark This revision replaces the previous versions, including the documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
22 June 2026

Certification Board