

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Shenzhen Kaifa Technology (Chengdu) Co., Ltd.  
No. 99 Tianquan Rd.  
Hi-Tech Development Zone  
Chengdu  
P.R.C.

Measuring instrument A static **Active Electrical Energy Meter**

Type : MA309M

Manufacturer's mark or name : Shenzhen Kaifa Technology (Chengdu)  
Co., Ltd.

Reference voltage : 3x57,7/100 to 3x240/415V

Reference current : 1 A or 5 A

Destined for the measurement of : electrical energy, in a  
- three-phase four-wire network

Accuracy class : B or C

Environment classes : M1 / E2

Temperature range : -40 °C / +75 °C

Further properties are described in the annexes  
– Description T13167 revision 0  
– Documentation folder T13167-1

Valid until 26 June 2036

Initially issued 26 June 2026

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
26 June 2026

## Certification Board