



EU-type examination certificate

Number **T13183** revision 0
Project number 4059876
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU,
after having established that the measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer

Conrad Electronic SE
Klaus-Conrad-Str. 1
D-92242 Hirschau
Germany

Measuring instrument

A static **Active Electrical Energy Meter**

Type : TC-DRS-205A MID, TC-DRS-205C (Modbus / M-Bus) MID, TC-DRS-205C CT (Modbus / M-Bus) MID

Manufacturer's mark or name : Conrad Electronic SE

Reference voltage : 230 V

Reference current : 5 A (Direct Connected variants)
1,5 A (Current Transformer operated variant)

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -25 °C / +55 °C
-40 °C / +70 °C (TC-DRS-205C CT, TC-DRS-205C CT Modbus MID, TC-DRS-205C CT M-Bus MID)

Further properties are described in the annexes:

- Description T13183 revision 0;
- Evaluation Report ER13183 revision 0;
- Documentation folder T13183-1.

Valid until

17 July 2036

Initially issued

17 July 2026

Issuing Authority

NMi Certin B.V., Notified Body number 0122
17 July 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)
ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

