EU-type examination certificate

Number **T13267** revision 3
Project number 3948724

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

EYKON AG
Alte Steinhauserstrasse 18
CH-6330 Cham
Switzerland

Authorised representative

EYKON Oy
Paperitehtaankatu 9
FI-40100 Jyväskylä
Finland

Measuring instrument

A static poly phase Active Electrical Energy Meter

Type	:	E360-AM3D...S3
Manufacturer's mark or name	:	EYKON AG
Reference voltage	:	- 3x230/400 V - 230 V
Reference current	:	5 A or 10 A
Destined for the measurement of	:	electrical energy, in a - three-phase four-wire network - single-phase two-wire network
Accuracy class	:	A or B
Environment classes	:	M1 / E2
Temperature range	:	-40 °C / +70 °C

Further properties are described in the annexes:
– Description T13267 revision 3;
– Documentation folder T13267-3;
– Evaluation report ER13267 revision 3.

Valid until

15 June 2036

Initially issued

15 June 2026

Remark

This revision replaces the earlier versions, except for its documentation folder

Issuing Authority

NMi Certin B.V., Notified Body number 0122
10 November 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at https://webgate.ec.europa.eu/single-market-compliance-space/home

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

RvA PRODUCTS RvA C 681