

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Tokheim Sofitam Applications S.A.S part of Dover Fueling Solutions  
Immeuble Le Cézanne  
Paris Nord, 31 - 35 Allée des Impressionnistes  
BP 45027 Villepinte  
95912 Roissy Ch de Gaulle Cedex  
France

Measuring instrument

**A fuel dispenser**

Manufacturer's mark or name : Tokheim Sofitam Applications S.A.S.  
part of Dover Fuelling Solutions.

Type : QUANTIUM "XXXX"\*

Accuracy class : 0,5

Destine for the measurement of : Gasoline; diesel

Further properties are described in the annexes:

- Description T10001 revision 51;
- Documentation folder T10001-8.

Valid until

30 October 2026

Remarks

- This revision replaced the previous revisions;
  - The documentation folder is not changed;
  - The fuel dispenser may be combined with one or more LPG- and/or gasoline oil dispensers and/or measuring systems for AdBlue, which make use of the same or separate calculating/indicating device(s);
  - The fuel dispenser may be combined with the self-service devices as mentioned in § 1.1 of the description.
- \* "XXXX" is a non-essential indication, see also § 1.3.1 the Description. This certificate covers the QUANTIUM-series, including the frame models Q100T, Q100T1, Q110, Q200T, Q200T MC, Q200T1, Q210, Q300T, Q300T1, Q310, Q400T, Q410, Q500T, Q500T1, Q510, etc.

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**  
24 August 2022

Certification Board

**NMi Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.