

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Emerson Process Management Flow B.V.
Neonstraat 1
6718 WX Ede
The Netherlands

Measuring instrument

An interruptible or non-interruptible **measuring system** for liquids other
than water

Manufacturer's mark or name : Emerson

Type designation : MMI-MID 001

Accuracy class : 0,3; 0,5; 1,0; 1,5; 2,5

Destined for the measurement of : Oil and oil products, alcohol,
potable liquids, chemicals,
liquefied gases under pressure and
cryogenic liquids.

Further properties are described in the annexes:

- Description T10071 revision 13;
- Documentation folder T10071-5.

Valid until

25 April 2027

Remarks

- This revision replaces the previous revisions;
- The documentation folder is not changed;
- The measuring instrument is approved for measuring mass, density at
metering conditions, density at base conditions, volume at metering
conditions and volume at base conditions and concentration;
- The measuring system can be fixed or movable.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
30 November 2020

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at
[http://ec.europa.eu/growth/tools-
databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally
signed and sealed. The digital
signature can be verified in the
blue ribbon at the top of the
electronic version of this
certificate.