

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Wayne Fueling Systems Sweden AB Hanögatan 10 211 24 MALMÖ Sweden
Measuring instrument	A fuel dispenser Manufacturer mark or name : Wayne Fueling Systems Sweden AB ^[1] Type : Global Ovation iX Accuracy class : 0,5 Environment classes : M1 / E1 Temperature range liquid : -25 °C – +50 °C Temperature range ambient : -25 °C – +55 °C; -40 °C – +55 °C for the electronics, in case a heater is applied. Minimum – maximum flow rate : 3 L/min – 80 L/min; 3 L/min – 130 L/min in case of two parallel connected gas separators and measurement transducers. Viscosity range : 0,4 – 8,0 mPa·s Further properties are described in the annexes: – Description T10085 revision 6; – Documentation folder T10085-1.
Valid until	3 April 2018
Remarks	– This revision replaces the earlier versions, except for its documentation folder.

^[1] Also see the remark in paragraph 1.3.2.

Issuing Authority

NMi Certin B.V., Notified Body number 0122

19 February 2015


C. Oosterman
Head Certification Board

NMi Certin BV
Hugo de Grootplein 1
3314 EG Dordrecht
PO Box 394
3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl)

Reproduction of the complete document is permitted.