



# EC-type examination certificate

Number **T10191** revision 0

Project number 811701

Page 1 of 1

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The "Metrologiewet" (Stb 2006, 137) as Dutch implementation of the Directive 2004/22/EC on measuring instruments (MID).

Applicant Endress + Hauser B.V.  
Nikkelstraat 6 – 12  
1411 AK Naarden  
The Netherlands

Submitted **An interruptible measuring instrument for filling drums.**  
Manufacturer : CSK Leeuwarden  
Designation : Vatenvuller

Characteristics Qmin – Qmax : 35 – 700 kg/min  
Minimum measured quantity : 20 kg  
Accuracy class : 0.5  
Environment class : M2 / E2  
Temperature range liquid : - 5 °C / +35 °C  
Temperature range ambient : -25 °C / +55 °C; condensing humidity;  
open location  
Intended for the measurement of : Stremsel

In the description T10191 revision 0 the additional characteristics are given.

Valid until 20 March 2019

Description and Documentation The measuring instrument is described in the description number T10191 revision 0 and documented in the documentation folder number T10191-1, appertaining to this EC-type examination certificate.

The Notified Body no. 0122  
NMI Certin, 20 March 2009

  
C. Oosterman  
Head Certification Board

NMI Certin BV  
Hugo de Grootplein 1  
3314 EG Dordrecht  
PO Box 394  
3300 AJ Dordrecht, NL  
T +31 78 6332 332  
F +31 78 6332 309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.  
Reproduction of the complete document is permitted.

