

## EC-type examination certificate

Number T10193 revision 0 Project number 804674 Page 1 of 1

NMi Certin B.V. Issued by

> Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation

2004/22/EC on measuring instruments (MID).

Netmill Co Ltd. Applicant

> Suite 38-40 London Fruit Exchange Brushfield Road London E1 6EU

United Kingdom

An interruptible measuring instrument for Truck loading In respect of

> Manufacturer Netmill Co Ltd.

MC5XX (XX see paragraph 1.2 of the description) Type

Characteristics Q<sub>min</sub> - Q<sub>max</sub> : See paragraph 1.2 of the description.

: 0.5 Accuracy class **Environment classes** : M3 / E2

Temperature range liquid : -25 °C / +55 °C : -25 °C / +55 °C Temperature range ambient

Intended for the measurement of Hydrocarbon based oils with maximum

viscosity of 20 mPa.s at 20°C.

In the description number

Valid until 14 February 2019

Description and The instrument is described in the description number T10193 revision 0

documentation and documented in the documentation folder T10193-1, appertaining to this

EC-type examination certificate.

An overview of the performed tests is given in Annex 1 belonging Remark

this Examination certificate.

Notified Body number 0122 NMi Certin, 23 June 2009

**Head Certification Board** 

NMi Certin BV Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 T+31 78 6332332

www.nmi.nl

3300 AJ Dordrecht, NL F+31 78 6332309 certin@nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

permitted.

